

A Determination of the Local Gravitational Acceleration for the Tsinghua Tabletop Kibble Balance

Weibo Liu, Nanjia Li, Yongchao Ma, Ruo Hu, Shuqing Wu, Wei Zhao, Songling Huang,
Senior Member, IEEE, Shisong Li[†], Senior Member, IEEE

Abstract—The Kibble balance requires a measurement of the local gravitational acceleration, g , with a typical relative measurement uncertainty of 10^{-9} . In this paper, the determination of g for the Tsinghua tabletop Kibble balance is presented. A polynomial fitting method is proposed for blind transfers of the absolute gravitational acceleration using relative gravimeters, showing agreement with the value obtained by the tide correction within a few parts in 10^9 . Horizontal and vertical gravity gradients are extracted by mapping the gravity distribution at different heights. The self-attraction effect of major components in the experiment, as well as some time-varying systematic effects, are modeled. The final determination of the gravitational acceleration at the mass position, with an uncertainty of $5.4 \mu\text{Gal}$ ($k = 2$), is achieved for the Tsinghua tabletop Kibble balance experiment.

Index Terms—Kibble balance, gravitational acceleration, absolute gravity measurement, relative gravity measurement.

NOMENCLATURE

AG	- Absolute gravimeter
RG	- Relative gravimeter
NIM	- National Institute of Metrology, China
THU	- Tsinghua University
SG	- Superconducting gravimeter
HGG	- Horizontal gravity gradient
VGG	- Vertical gravity gradient
FEA	- Finite Element Analysis
IERS	- International Earth Rotation and Reference System Service
Δg_s	- Gravity correction of instrument self-attraction
Δg_t	- Time-varying gravity correction
Δg_{ta}	- Gravity correction of atmospheric mass effect
Δg_{tp}	- Gravity correction due to the polar motion
Δg_{tt}	- Gravity correction of the Earth's tides
Δg_{HGG}	- Gravity correction related to horizontal field distribution

Δg_{VGG}	- Gravity correction related to vertical field distribution
g_r	- Absolute gravitational acceleration at the reference site
g_0	- Absolute gravitational acceleration at the reference point of the Kibble balance site
g_m	- Absolute gravitational acceleration at the mass position of the Kibble balance site
g	- Absolute local gravitational acceleration
m	- Test mass to be calibrated by Kibble balance
Bl	- Magnetic geometrical factor of a Kibble balance
I	- Current through the coil in the weighing phase
U	- Coil induced voltage in the velocity phase
v	- Coil moving velocity in the velocity phase
O	- Polynomial fit order
τ	- Time delay of tides in two sites
Δg	- Gravity offset of two sites/measurements
H	- Height of the horizontal gravity mapping
h_0	- Height of the CG6 sensor to the ground surface
H_{lab}	- Height of the Kibble balance laboratory
p	- Pressure (in hPa) at the mass position
ω	- Angular velocity of the Earth's rotation
R	- Radius of the Earth
λ, ϕ	- Geodetic coordinates of the measurement point
x, y	- Pole coordinates updated by the IERS
k_{op}	- Gain factor of the Tsoft tide estimation
Δg_o	- Offset of the Tsoft tide estimation
z_w	- Distance between the bottom of the vacuum chamber and the mass position

I. INTRODUCTION

THE KIBBLE balance, an instrument originally proposed by Dr. Bran Kibble [1] offering a precision link between a test mass m and the Planck constant h , stands as one of the primary methods for mass realizations at the kilogram level and beyond. Presently, numerous metrology institutes are actively engaged in Kibble balance experiments, exemplified by initiatives from institutions such as [2]–[13].

A conventional Kibble balance comprises two distinctive measurement phases: the weighing phase and the velocity phase. In the weighing phase, the gravitational force acting on the test mass, mg , is counterbalanced by an electromagnetic force generated by a current-carrying coil in a magnetic field, expressed as $mg = BLI$, where I denotes

W. Liu, N. Li, Y. Ma, W. Zhao, S. Huang, and S. Li are with the Department of Electrical Engineering, Tsinghua University, Beijing 100084, China.

R. Hu and S. Wu are with the National Institute of Metrology (NIM), Beijing 100029, China.

This work was supported by the National Key Research and Development Program under Grant 2022YFF0708600 and the National Natural Science Foundation of China under Grant 52377011.

[†]Email: shisongli@tsinghua.edu.cn

the current in the coil and Bl represents the dot product of magnetic flux density and current flowing path, so-called the magnetic geometrical factor. Subsequently, in the velocity phase, the Bl term is calibrated by moving the coil with a velocity v in the same magnetic field, resulting in an induced voltage on coil terminals $U = Blv$. With Bl eliminated, the mass is determined as $m = \frac{UI}{gv}$. Details of the Kibble balance principle can be found in recent review papers, e.g. [14]. The most accurate Kibble balance can calibrate mass m with a relative uncertainty of about one part in 10^8 [2], hence necessitating a determination of the local gravitational acceleration g at the order of 10^{-9} , e.g. [5], [15]–[22].

In late 2022, Tsinghua University initiated a tabletop Kibble balance project with the objective of creating an accurate, robust, compact, and open-hardware mass realization instrument [23], [24]. The goal of the measurement uncertainty is below $50 \mu\text{g}$ for mass calibrations ranging from 10 g to 1 kg. Progress in the Tsinghua tabletop Kibble balance measurement system has recently transpired. Herein, we report the determination of the local gravitational acceleration, g , for the Tsinghua tabletop Kibble balance. A generalized procedure for g measurement is presented, accompanied by a detailed discussion of the measurement specifics.

Note that to be consistent with the gravity community, the non-SI unit, Gal, is used in the following manuscript, where $1 \text{ Gal} = 1 \times 10^{-2} \text{ m/s}^2 \approx 10^{-3} g$, $1 \text{ mGal} = 1 \times 10^{-5} \text{ m/s}^2 \approx 10^{-6} g$ and $1 \mu\text{Gal} = 1 \times 10^{-8} \text{ m/s}^2 \approx 1 \times 10^{-9} g$.

II. GENERALIZED g MEASUREMENT PRINCIPLES

Fig. 1 illustrates a generalized schematic flow for determining the local gravitational acceleration (g) in Kibble balance experiments. The process involves three key steps, elaborated upon below.

- 1) *Determine the absolute gravitational acceleration g_0 at a reference position in the Kibble balance site.*

In this step, two general methods are typically employed. The first involves the setup of an absolute gravimeter, such as classic free-falling interferometry [25] or an atomic gravimeter [26], at the reference position of the Kibble balance room. This allows for the direct measurement of the absolute value of g_0 . The second method entails transferring the absolute gravitational acceleration value, g_r , from a reference site (absolute g value is precisely known) using relative gravimeters. In this approach, it is crucial to account for time-varying corrections during the measurement, particularly those related to the Earth's tides. Typically, repeated ABA or ABBA measurements are required to reduce the Type A uncertainty and the Type B uncertainty components caused by the instrument drift. Performing more measurement loops in a short period of time helps to reduce both types of uncertainties, and the number of measurements at each site, N depends on




Fig. 1. A schematic overview for the absolute determination of gravitational acceleration (g) in Kibble balances is presented. In the diagram, AG and RG represent absolute and relative gravimeters, respectively. g_r denotes the value at an external reference site, while g_0 represents the value at the reference site within the Kibble balance room. VGG and HGG depict the vertical and horizontal gravity gradients, respectively. g_m signifies the transferred value at the mass position. Corrections, primarily related to tides, are encapsulated in Δg_t , and Δg_s accounts for the correction stemming from the instrument's self-attraction. The algorithms for g calculations are denoted as 1, 2, and 3.

the best overall uncertainty required. For example, if a CG6 relative gravimeter (standard deviation $<5 \mu\text{Gal}$, compensated daily drift $<20 \mu\text{Gal}$) is used, N greater than 6 can result in an instrument-related uncertainty of less than $2 \mu\text{Gal}$.

- 2) *Determine the absolute gravitational acceleration g_m at the mass position in Kibble balances.*

In Kibble balances, the gravitational acceleration at the mass location, g_m , is required. Therefore, a gravity transfer from the reference position of the Kibble balance to the mass position ($g_0 \rightarrow g_m$) should be performed. For this purpose, a mapping of the gravity gradient in both vertical and horizontal directions (VGG and HGG) must be established. The conventional approach is to measure the gravity change at various known horizontal and vertical positions using relative gravimeters and then apply a fitting algorithm to determine the difference of the gravitational acceleration, $g_m - g_0$.

- 3) *Determine the final gravitational acceleration g considering time-varying and self-attraction corrections.*

The main errors in the gravitational acceleration g obtained by the above steps include Δg_t , time-varying corrections, mainly the Earth's tides, and Δg_s , the correction from the self-attraction of the instrument. The tides lead to corrections depending on the time of day,




Fig. 2. (a) describes the geographical positions of the NIM and THU laboratories, accompanied by the transmission path of the CG6 throughout the measurements. (b) demonstrates the arrangement of 24-point locations in the Tsinghua lab, where the red points serve as the reference point.

the latitude, the longitude, and the altitude. The tide correction can be measured by a relative gravimeter or using theoretical estimation tools such as Tsoft [27], ETERNA [28], Tamura tidal potential [29], etc. The instrument's self-attraction can be corrected by direction measurement or finite element analysis (FEA) calculations following the similarity between the law of universal gravitation and Coulomb's law.

In the following sections, the determination of g for the Tsinghua tabletop Kibble balance following the above three steps is presented.

III. g_0 DETERMINATION

For the Tsinghua Kibble balance, the relative gravity transfer approach ($g_r \rightarrow g_0$) is employed to determine g_0 at the Kibble balance site. In Fig. 2 (a), the reference site providing the g_r value is a point utilized in the 2017 Key Comparison of Absolute Gravimeters (CCM.G-K2.2017) [30], situated at the Gravimetric building, Changping campus of the National Institute of Metrology (NIM, China). The g_r value is recorded as 980 122 922.8 μGal with a standard uncertainty of 1.0 μGal .

The Tsinghua tabletop Kibble balance is situated in a laboratory on the fourth floor of the west main building, Tsinghua University. This distinctive Kibble balance is installed within a building rather than on the Earth's surface or in an underground base, with the purpose of testing its resilience in a general-conditioned laboratory environment. Fig. 2 (b) depicts the top view of the measurement site, with the Tsinghua tabletop Kibble balance anticipated to be installed near the center of the 2.5 m $\times$ 1.9 m room. A gravity reference point, denoted by the red cross in Fig. 2(b), is chosen as the g transferred point. Positioned approximately 0.85 m away from the vacuum chamber, this allows space for future absolute g checks or relative gravity monitoring.

As illustrated in Fig. 2 (a), the linear distance between the reference gravity site (NIM) and the Tsinghua Kibble balance site (THU) is approximately 28.7 km. To measure the gravity difference between NIM and THU sites, a relative gravimeter CG6 (Scintrex) [31], a spring-type relative gravimeter with a measurement range of 8000 mGal, resolution of 0.1 μGal , standard deviation $< 5 \mu\text{Gal}$, and compensated daily drift $< 20 \mu\text{Gal}$, is utilized. Following the transportation route (blue line, approximately 33 km drive), the CG6 is transferred between the two sites using a vehicle, alternately measuring the gravitational acceleration. The measurement, lasting about 14 hours, involves seven loops. Implementing the ABA measurement strategy allows the removal of the linear drift of the instrument, enabling precise determination of the gravity difference.

Given the geographical separation of the two measurement sites, variations in gravity corrections due to the Earth's tides may impact the measurement results differently. In consideration of this discrepancy and to streamline the measurement process, the tide correction within the CG6 is disabled. To account for tidal influences, Tsoft [27], a program incorporating date, latitude, longitude, and altitude inputs to calculate tidal effects on gravity, is employed. Independent calculations of tidal influence, presented in Fig. 3 (a) and (b), are conducted for NIM and Tsinghua, respectively.

As illustrated, there is a difference of approximately -0.5 μGal in the mean value and a 1 μGal peak-to-peak value between the tidal influences, attributed to a phase shift (time delay). To validate the accuracy of the tide estimation by Tsoft, an experimental measurement using a superconducting relative gravimeter iGrav-SG (GWR Instruments) at the NIM site [21] was conducted during the g transfer period. The iGrav-SG [32] has a precision of $10^{-3} \mu\text{Gal}$ in the frequency domain and 0.05 μGal in the time domain for one-minute averaging, and can be used as an ultra-high precision continuous gravity reference for measurements. As depicted in Fig. 3 (b), the maximum difference between the experimental and Tsoft estimation is within 2 μGal .

The seven-loop CG6 measurement results at the NIM and THU sites are presented in Fig. 3 (c) and (d). Following the tide removal process, the mean g measurement at the NIM site is 4 661.371 9 mGal, with a standard deviation




Fig. 3. (a) illustrates the tidal estimations at the NIM and THU sites, alongside the relative g measurements at the NIM site using the superconducting gravimeter iGrav-SG. In (b), the disparities between the tide estimations and measurements are presented. (c) and (d) showcase the CG6 measurement outcomes at NIM and THU, respectively. The dashed lines represent the tides estimated by Tsot. The right side of (c) and (d) displays the average gravity values after the tide correction.

of $4.2 \mu\text{Gal}$. Similarly, at the THU Kibble balance site, the mean measurement is 4678.0303 mGal , with a standard deviation of $3.0 \mu\text{Gal}$. The slightly lower data deviation at the THU site indicates that the vibration noise floor of the Tsinghua Kibble balance laboratory may be low, at least in the sensitive range of the CG6. This could be attributed to the proximity of the laboratory to the building mass center. The gravity difference between the THU site and the NIM site is calculated as 16.6584 mGal , with a standard uncertainty of $2.1 \mu\text{Gal}$. Using the NIM reference value, the gravitational acceleration at the reference point in the Tsinghua tabletop Kibble Balance room is determined as $g_0 = (980\,139\,581.2 \pm 2.3) \mu\text{Gal}$ ($k = 1$).

During the data processing, we observed that if the tide difference between the reference site (NIM) and the measurement site (THU) is negligible, a polynomial fit method can estimate the gravity difference without prior knowledge of tide information. The concept is to assume that over a short period, such as 12 hours, the tide correction's shape can be adequately represented by a polynomial fit of order O , e.g., $O = 6$. This assumption can be validated using data from the superconducting gravimeter iGrav-SG. Fig. 4 displays the residuals of a 12-hour tide measurement obtained by the superconducting gravimeter iGrav-SG and polynomial fittings of various orders ($O = 4, 5, 6, 7$). Notably, for $O \geq 5$, the tide difference




Fig. 4. (a) presents a 12-hour tide measurement result using the superconducting gravimeter iGrav-SG. (b) shows the differences between the experimental and the O -order polynomial fittings. The fitting orders, $O = 4$ to $O = 7$, are shown in the plot.

between the experimental data and the fit remains within $\pm 1 \mu\text{Gal}$.

With a fixed-order polynomial fit representing the tides, the g transfer can be simplified as

$$\begin{aligned} & \text{Min}\{\text{Res.}\{\text{Polyfit}\{g_1(t), g_2(t - \tau) + \Delta g\}\}\} \\ & \text{s.t. } \tau = \text{const.}, \Delta g = \text{const.} \end{aligned} \quad (1)$$

where $g_1(t)$ and $g_2(t)$ are the time series of the relative g measurement at two sites, τ (time delay of tides in two sites) and Δg (gravity offset) constant parameters to be determined. The goal of the estimation is to make the lowest residual of the polynomial fit.

Here we take the CG6 data as an example and set the fit order $O = 6$. g_1 and g_2 are respectively the measurement results at the NIM and THU sites. The standard deviation of the residual, σ , is set as the target for minimization. As shown in Fig. 5, by scanning τ and Δg , the optimal parameters, $\tau_{\text{op}} = 15 \text{ s}$ and $\Delta g_{\text{op}} = 16.6576 \text{ mGal}$, are obtained. In this case, the standard deviations of the NIM measurement and THU measurement are respectively $2.5 \mu\text{Gal}$ and $1.5 \mu\text{Gal}$. This blind determination of the gravitational acceleration at the reference point in the Tsinghua tabletop Kibble Balance room yields $g_0 = (980\,139\,580.4 \pm 1.6) \mu\text{Gal}$ ($k = 1$). The result agrees well with the value obtained by the conventional tide correction and has a slightly lower measurement uncertainty.

IV. g_m DETERMINATION

The preceding section establishes the absolute gravity value at a designated reference point within the Tsinghua Kibble balance room. In order to ascertain the gravity value at the specific experimental mass position, located approximately 0.85 m horizontally and 1 m vertically from the reference point on the floor, it is imperative to determine both the horizontal gravity gradient (HGG) and vertical gravity gradient (VGG). This necessitates the mapping of the gravity distribution across the experiment room.

As presented in Fig. 2 (b), the floor of the Tsinghua Kibble balance laboratory is segmented into a grid arrangement of 24 points, organized in four rows and six columns, labeled from 1 to 24, with point 5 serving as the




Fig. 5. (a) plots the standard deviation of the residual for the polynomial fit at different gravity offsets Δg and time delay τ . (b) illustrates the best polynomial fit curve and data points, while (c) presents the residuals of the best fit.




Fig. 6. Relative gravity measurement results at different heights ($H = 0\text{ m}$, 0.3 m , 0.6 m). For each group of data, from left to right correspond points 1 to 24.

reference position. The gravitational acceleration (g) values at three distinct height levels, i.e. $H = 0\text{ m}$, $H = 0.3\text{ m}$, and $H = 0.6\text{ m}$, are systematically mapped using the CG6 instrument. The outcomes of these measurements, segregated by height levels, are presented in Fig. 6. By subtracting the tidal influence from the measured data, it becomes feasible to derive the relative gravity distribution in three-dimensional space.

A. HGG

The Kibble balance room is situated in the central area of the west main building at Tsinghua University,




Fig. 7. The upper subplots show, from left to right, the change in g at ground level, 0.3 m above ground level, and 0.6 m above ground level. The unit of the contour is μGal . The lower subplot is an average of the top three plots. The red circle, KB, is where the Tsinghua Kibble balance will be installed.

characterized by a flat terrain with an altitude of approximately 64 m above sea level on the ground floor. Given the relatively confined dimensions of the Kibble balance experimental space ($1.9\text{ m} \times 2.5\text{ m}$), it is anticipated that the horizontal gradient (HGG) is insignificantly small. The top subplots of Fig. 7 depict the relative gravitational distribution at three distinct heights, $H = 0\text{ m}$, 0.3 m , and 0.6 m . Note that the relative g value at the reference position (point 5) is normalized to zero, and cubic interpolation is employed to enhance the smoothness of the plots. The lower subplot represents the average of the three plots. The designated location for the installation of the Tsinghua Kibble balance is denoted by the red circle labeled 'KB'.

Despite inherent measurement uncertainties, it is challenging to discern a distinct horizontal gravity gradient along the horizontal axes. To simplify the analysis, the mean of the mapping within the central rectangle (formed by points 10-11-15-14) is employed to quantify the g difference between the reference point and the experimental mass position. This yields differences of $(1.2 \pm 1.5)\mu\text{Gal}$, $(-0.7 \pm 2.7)\mu\text{Gal}$, and $(3.2 \pm 0.9)\mu\text{Gal}$ for $H = 0\text{ m}$, $H = 0.3\text{ m}$, and $H = 0.6\text{ m}$, respectively. The weighted mean of the three measurements is $(2.4 \pm 0.7)\mu\text{Gal}$. A cross-check directly using the measurement data at points 10, 11, 14, and 15 is conducted, yielding an average of $(2.2 \pm 1.2)\mu\text{Gal}$, which agrees well with the value obtained by fittings. Notably, no substantial variation in HGG is observed as a function of height (H).

To further scrutinize the HGG, the gravity maps at the three distinct levels are averaged and presented in the




Fig. 8. The upper subplot displays the curve fitted for the vertical gravity gradient. The lower subplot depicts the residuals at varying heights.

lower subplot of Fig. 7. It can be seen that the gravity distribution remains within a few μGal , with no discernible global directional gradient.

B. VGG

The data presented in Fig. 6 illustrates a linear change in g along the vertical direction. To assess this linearity, linear fits were conducted using three distinct height g data points for each location, and the average of all 24 fitted curves is depicted in Fig. 8 (a). The resulting ultimate curve from this fitting process is characterized by $\partial g/\partial H = -296.2 \mu\text{Gal}/\text{m}$, showcasing an exceptionally linear fit with $R^2 \approx 0.99994$ and residuals below $1 \mu\text{Gal}$, as illustrated in Fig. 8 (b).

Given that g changes by $-2.962 \mu\text{Gal}$ per centimeter vertically, precise measurement of the vertical distance between the mass position and the CG6 sensor at the reference point becomes crucial. Considering the sensor height of the CG6, the g value at the mass position can be expressed as

$$g_m = g_0 + \Delta g_{\text{HGG}} + \Delta g_{\text{VGG}}, \quad (2)$$

where $\Delta g_{\text{VGG}} = (\partial g/\partial H)(H - h_0)$; H represents the height of the mass position from the floor surface, and h_0 denotes the height of the gravity measurement sensor. For the CG6 experimental setup here, h_0 includes the height of a supporting platform and is approximately 21.2 cm.

V. FINAL RESULT OF g

A. Time-varying corrections, Δg_t

In our analysis, we consider three significant time-varying corrections: the atmospheric mass effect, the Earth's tides, and the polar motion effect.

The atmospheric mass effect is estimated in accordance with the US Standard Atmosphere 1976 [33]. The atmospheric pressure p (in hPa) is measured at the laboratory, and the correction for gravitational acceleration is expressed as

$$\Delta g_{\text{ta}} = -0.3 \left[p - 1013.25 \left(1 - \frac{0.0065 H_{\text{lab}}}{288.15} \right)^{5.2559} \right], \quad (3)$$

TABLE I
Tsoft PARAMETERS USED FOR ESTIMATION THE EARTH'S TIDES.

Group name	Min freq	Max freq	Amplitude factor	phase shift
DC	0.000000	0.000001	1.00000	0.0000
long	0.000140	0.002427	1.15800	0.0000
Mf	0.002428	0.249951	1.15738	0.0000
Q1	0.721500	0.906315	1.15424	0.0000
O1	0.921941	0.940487	1.15423	0.0000
P1	0.958085	0.998028	1.14916	0.0000
K1	0.999853	1.003651	1.13493	0.0000
PSI1	1.005329	1.005623	1.26951	0.0000
PHI1	1.007595	1.011099	1.17025	0.0000
OO1	1.013689	1.216397	1.15626	0.0000
All2	1.719381	2.182843	1.16188	0.0000
M3	2.753244	3.381478	1.07352	0.0000
M4	3.381379	4.347615	1.03900	0.0000

where H_{lab} is the elevation of the measurement point. For the Tsinghua tabletop Kibble balance, H_{lab} is approximately 64 m. The standard uncertainty assigned for the atmospheric mass correction is $0.5 \mu\text{Gal}$.

The gravitational field change due to the polar motion can also be corrected using a formula [16], i.e.

$$\Delta g_{\text{tp}} = 1.16 \omega^2 R \sin 2\phi (x \cos \lambda - y \sin \lambda) \times 10^8, \quad (4)$$

where ω is the angular velocity of the Earth's rotation, R is the radius of the Earth, λ and ϕ represent the geodetic coordinates of the measurement point, and x , y are the pole coordinates updated by the International Earth Rotation and Reference System Service (IERS). The unit of the Δg_{tp} correction is μGal , and the standard uncertainty assigned to the polar motion correction is $0.1 \mu\text{Gal}$.

The Earth's tide correction constitutes the most substantial time-varying component in g measurement, with a maximum tide variation of up to $280 \mu\text{Gal}$ or approximately 2.8×10^{-7} . For the tide correction, Tsoft is employed to estimate the changes in tides over time, and the parameters used are detailed in Tab. I.

To ensure the accuracy of Tsoft estimation, a calibration procedure is implemented using the superconducting gravimeter iGrav-SG. The upper subplot of Fig. 9 illustrates a two-day measurement result from both the superconducting gravimeter iGrav-SG and the Tsoft estimation. The calibration aims to determine the optimal gain factor k by minimizing the standard deviation of $g(\text{SG}) - k \cdot g(\text{Tsoft})$. Through numerical search, the optimal gain factor is found to be $k_{\text{op}} = 1.008$. The lower subplot of Fig. 9 demonstrates the residual between the superconducting gravimeter iGrav-SG measurement and the Tsoft estimation in this case. An offset, $\Delta g_o = (-14.4 \pm 0.8) \mu\text{Gal}$, is identified.

In summary, the Earth's tide correction is expressed as

$$\Delta g_{\text{tt}} = k_{\text{op}} \cdot g(\text{Tsoft}) + \Delta g_o. \quad (5)$$

The standard uncertainty associated with the Earth's tide correction is $0.8 \mu\text{Gal}$.




Fig. 9. The upper plot shows the tides comparison of a two-day measurement from the superconducting gravimeter iGrav-SG and the Tsoft estimation. The lower plot is the difference between two results after the gain k_{op} is multiplied, i.e. $g(\text{SG}) - k_{op} \cdot g(\text{Tsoft})$. The dashed lines show the boundaries of the standard deviation of the data.

B. Self-attraction correction, Δg_s

The mass distribution within the Kibble balance instrument induces a gravitational field, thereby necessitating a self-attraction correction for g . The self-attraction effect in a conventional Kibble balance typically exhibits an impact on the order of a few parts in 10^8 . Given the g measurement relative uncertainty target of 10^{-9} , a determination of this effect with a 10% uncertainty is deemed sufficient. At this precision level, three methods are viable: 1) direct calculation of the effect based on Newton's law of gravitation, 2) measurement of the relative gravity change using gravimeters, and 3) finite element analysis (FEA) calculations utilizing the analogy between Coulomb's law and Newton's law of gravitation. For tabletop Kibble balances, direct measurement proves challenging due to spatial constraints. Consequently, FEA calculations are employed in this context.

The self-attraction field generated by a focused mechanical segment is written as

$$\Delta g_s = G \frac{\rho V}{|r|^3} \vec{r} \cdot \vec{k}, \quad (6)$$

where G is the gravitational constant, ρ and V respectively the density and volume of the segment, $\vec{r}$ the vector from the mass position to the segment mass center, $\vec{k}$ the unit vector along the vertical direction z . When the segment is imported into the FEA package, it needs to first change the material into the vacuum and then set a constant charge density numerically $\rho' = \rho$. Finally, the vertical electrical field at the mass position E_z can be solved by FEA following

$$E_z = \frac{1}{4\pi\epsilon_0} \frac{\rho' V}{|r|^3} \vec{r} \cdot \vec{k}, \quad (7)$$

where ϵ_0 is the vacuum permittivity. Comparing (6) and (7), the self-attraction is determined as

$$\Delta g_s = 4\pi\epsilon_0 G E_z. \quad (8)$$

The CAD model of Tsinghua tabletop Kibble balance is shown in Fig. 10(a). In the calculation, only a few major components, including the vacuum chamber, the balance major framework, and the magnet, are considered. The




Fig. 10. (a) shows the CAD model of the Tsinghua tabletop Kibble balance. (b) present the self-attraction effect of major mechanical pieces as a function of vertical distance. Note that the zero position $z = 0$ is set at the bottom of the vacuum chamber. (c) plots the gravitational field map at the weighing horizontal plane.

self-attractive field, g_s , along the vertical direction, is shown in Fig. 10(b). The reference position, $z = 0$, is chosen at the bottom of the vacuum chamber and the mass position is at $z_w = 0.918$ m. Note that z_w contains half of the height of a stainless steel mass (64 mm/2=32 mm). It can be seen that the relative uncertainty caused by the self-attraction at the mass position is well below 1×10^{-8} . The contribution of the chamber and the balance (including the magnet) is respectively $-0.2 \mu\text{Gal}$ and $2.2 \mu\text{Gal}$. In Fig. 10(c), the self-attraction distribution at the horizontal




Fig. 11. A cause-and-effect graph of the value and the uncertainty components up to the determination of the final g value. Here Σ denotes that the output quantity equals the sum of the inputs, while W means the weighted mean.

plane of the mass position (stainless steel mass) is mapped. The self-attraction value varies slightly inside the vacuum chamber, mainly due to the near-field mass distribution.

A specific feature of the Tsinghua Kibble balance is to move the magnet instead of the coil during the velocity measurement. The movement of the magnet may reshape the self-attraction curve along the vertical. However, the gravitational field change is about $0.3 \mu\text{Gal}$ over a 10 mm measurement range, which compared to the measurement uncertainty target is negligible.

The uncertainty of self-attraction correction mainly depends on the material density and segment dimensions, since the calculation accuracy is, in general, well below 10%, a standard uncertainty of $0.5 \mu\text{Gal}$ is assigned.

C. Final result of g determination

With all the measurement and analysis, we can finally summarize the g determination result. The measurement uncertainty budget is presented in Tab. II, including the weighted mean of uncertainties from different methods, and merging different uncertainty components. A cause-and-effect graph of the value and the uncertainty components up to the determination of the final g value is also given in Fig. 11. The total g measurement uncertainty obtained is $5.4 \mu\text{Gal}$ ($k = 2$).

Specifically, the final g value is the sum of g_m , Δg_t and Δg_s , therefore, the total g measurement uncertainty is obtained by the square law using the uncertainties of g_m , Δg_t and Δg_s , i.e. $u_g = \sqrt{(u_{g_m})^2 + (u_{\Delta g_t})^2 + (u_{\Delta g_s})^2}$. The uncertainties of g_0 , g_m , and Δg_t can be similarly determined. For $(g_0 - g_r)$ and Δg_{HGG} , the average value and the uncertainty are obtained by the weighted mean of results corresponding to the different measurement methods. The weight for each method, w_i , is set as $w_i = 1/u_i^2$, where u_i is the measurement uncertainty for the i -th method.

VI. CONCLUSION

This paper presents the determination of the local gravitational acceleration, g , for the Tsinghua tabletop Kibble balance with a targeted relative uncertainty of 10^{-9} . The gravity transfer between an international gravity comparison site at NIM Changping campus and the Kibble balance laboratory at Tsinghua University has been carried out, alternately realized by using a relative gravimeter CG6. Beyond the conventional tide correction, a polynomial fitting method was introduced for the blind transfer of absolute gravitational acceleration. The results obtained through this approach exhibit great convenience with a good agreement with values derived from the tide correction. Furthermore, the investigation involved the extraction of horizontal and vertical gravity gradients by mapping the gravity distribution at various heights, providing valuable insights into the spatial variations of gravitational acceleration. To enhance the precision of the Kibble balance experiment, comprehensive modeling of time-varying effects and the self-attraction effect was conducted. The culmination of these efforts resulted in the achievement of the final determination of the gravitational acceleration at the mass position for the Tsinghua tabletop Kibble balance experiment. The reported uncertainty stands at $5.4 \mu\text{Gal}$ ($k = 2$), emphasizing a base for future precise realization of masses at the 10^{-8} level.

Looking ahead, two key tasks are planned: a direct g value check using absolute gravimeters and the vibration measurement. Once the experiment is fully set up, the direct g measurement will be conducted to complement and validate the g transfer method used in this study. Additionally, since the Tsinghua tabletop Kibble balance will be installed on the fourth floor of the building, addressing the effects of mechanical vibrations transmitted through the ground is crucial, particularly along the vertical axis. Vibration sensors will be installed at various locations within the building and inside the experiment to ensure accurate and reliable measurements.

ACKNOWLEDGEMENT

The authors would like to thank the Uber driver, Mr. Liu, for the 14-hour non-stop delivery of the CG6 instrument between the Tsinghua University site and the NIM Changping campus site. Thanks also to Zhilan Huang at Tsinghua University for organizing the gravity transfer. Shisong Li would like to thank his BIPM colleagues for the valuable discussions on the g measurement.

REFERENCES

- [1] B. P. Kibble, "A measurement of the gyromagnetic ratio of the proton by the strong field method," in *Atomic masses and fundamental constants 5*. Springer, 1976, pp. 545–551.
- [2] B. M. Wood, C. A. Sanchez, R. G. Green, and J. O. Liard, "A summary of the Planck constant determinations using the NRC Kibble balance," *Metrologia*, vol. 54, no. 3, pp. 399–409, 2017.
- [3] D. Haddad, F. Seifert, L. S. Chao, *et al.*, "Measurement of the Planck constant at the National Institute of Standards and Technology from 2015 to 2017," *Metrologia*, vol. 54, no. 5, pp. 633–641, 2017.

TABLE II
SUMMARY OF THE g DETERMINATION

	method	value/ μGal	uncertainty ($k = 2$)/ μGal
g_r	key comparison	980 122 922.8	2.0
$g_0 - g_r$	tide correction	16 658.4	4.2
	polynomial fit ($O = 6$)	16 657.6	2.4
	weighted mean	16 657.7	2.0
g_0	$g_r + (g_0 - g_r)$	980 139 580.5	2.8
Δg_{HGG}	gravity mapping($z = 0$)	1.2	3.0
	gravity mapping($z = 0.3$ m)	-0.7	5.4
	gravity mapping($z = 0.6$ m)	3.2	1.8
	weighted mean	2.4	1.4
Δg_{VGG}	gravity mapping	-296.2($H_m/m-0.212$)	3.6
g_m	$g_0 + \Delta g_{\text{HGG}} + \Delta g_{\text{VGG}}$	-	4.8
Δg_{ta}	US Standard Atmosphere 1976	-	1.0
Δg_{tp}	Recommended formula	-	0.2
Δg_{tt}	Tsoft estimation	-	1.6
Δg_t	$\Delta g_{\text{ta}} + \Delta g_{\text{tp}} + \Delta g_{\text{tt}}$	-	2.0
Δg_s	FEA, inverse square laws	2.0	1.0
g	$g_m + \Delta g_t + \Delta g_s$	-	5.4

- [4] L. Chao, F. Seifert, D. Haddad, J. Pratt, D. Newell, and S. Schlamminger, "The performance of the KIBB-g1 tabletop Kibble balance at NIST," *Metrologia*, vol. 57, no. 3, p. 035014, 2020.
- [5] A. Eichenberger, H. Baumann, A. Mortara, *et al.* "First realisation of the kilogram with the METAS Kibble balance," *Metrologia*, vol. 59, no. 2, p. 025008, 2022.
- [6] H. Fang, F. Bielsa, S. Li, A. Kiss, and M. Stock, "The BIPM Kibble balance for realizing the kilogram definition," *Metrologia*, vol. 57, p. 045009, 2020.
- [7] M. Thomas, D. Ziane, P. Pinot, R. Karcher, A. Imanaliev, F. P. Dos Santos, S. Merlet, F. Piquemal, and P. Espel, "A determination of the Planck constant using the LNE kibble balance in air," *Metrologia*, vol. 54, no. 4, pp. 468–480, 2017.
- [8] R. J. Hawke, M. T. Clarkson, "Position control for the MSL Kibble balance coil using a syringe pump," *ACTA IMEKO*, vol. 11, no. 4, pp. 1-7, 2022.
- [9] Z. Li, Y. Bai, Y. Wang, *et al.*, "Improvements of the NIM-2 Joule Balance Since 2020," *IEEE Transactions on Instrumentation and Measurement*, vol. 72, pp. 1–7, 2023.
- [10] D. Kim, M. Kim, M. Seo, *et al.*, "Realization of the kilogram using the KRISS Kibble balance," *Metrologia*, vol. 57, no. 5, p. 055006, 2020.
- [11] H. Ahmedov, R. Orhan, and B. Korutlu, "UME Kibble balance operating in air," *Metrologia*, vol. 60, no. 1, p. 015003, 2022.
- [12] I. A. Robinson, "The architecture of the NPL next generation Kibble balance," in *IMEKO 24th TC3, 14th TC5, 6th TC16 and 5th TC22 International Conference*, Cavtat-Dubrovnik, Croatia, 2022.
- [13] C. Rothleitner, I. Poroskun, S. Svitlov, *et al.*, "A method for evaluating the force factor in oscillating Kibble balances with application to the Planck-Balance," *Metrologia*, vol. 60, no. 5, p. 055007, 2023.
- [14] I. A. Robinson and S. Schlamminger, "The watt or Kibble balance: a technique for implementing the new SI definition of the unit of mass," *Metrologia*, vol. 53, no. 5, pp. A46–A74, 2016.
- [15] J. Liard, C. Sanchez, B. Wood, A. Inglis, and R. Silliker, "Gravimetry for watt balance measurements," *Metrologia*, vol. 51, no. 2, p. S32, 2014.
- [16] Z. Jiang, V. Pálinkáš, O. Francis, H. Baumann, J. Mäkinen, L. Vitushkin, S. Merlet, L. Tisserand, P. Jousset, C. Rothleitner *et al.*, "On the gravimetric contribution to watt balance experiments," *Metrologia*, vol. 50, no. 5, p. 452, 2013.
- [17] I.-M. Choi, K.-C. Lee, S. Lee, D. Kim, and H.-Y. Lee, "Gravity measurement for the KRISS watt balance," *IEEE Transactions on Instrumentation and Measurement*, vol. 66, no. 6, pp. 1317–1322, 2017.
- [18] S. Merlet, A. Kopaev, M. Diamant, G. Geneves, A. Landragin, and F. P. Dos Santos, "Micro-gravity investigations for the LNE watt balance project," *Metrologia*, vol. 45, no. 3, p. 265, 2008.
- [19] E. J. Leaman, D. Haddad, F. Seifert, L. S. Chao, A. Cao, J. R. Pratt, S. Schlamminger, and D. B. Newell, "A determination of the local acceleration of gravity for the NIST-4 watt balance," *IEEE Transactions on Instrumentation and Measurement*, vol. 64, no. 6, pp. 1663–1669, 2015.
- [20] A. Louchet-Chauvet, S. Merlet, Q. Bodart, A. Landragin, F. P. Dos Santos, H. Baumann, G. d'Agostino, and C. Origlia, "Comparison of 3 absolute gravimeters based on different methods for the e-Mass project," *IEEE Transactions on Instrumentation and Measurement*, vol. 60, no. 7, pp. 2527–2532, 2011.
- [21] J. Xu, J. Feng, Q. Wang, L. Mou, Z. Li, B. Han, L. Qian, R. Hu, S. Wu, Q. He *et al.*, "The determination of gravitational acceleration in the joule balance at NIM," *Metrologia*, vol. 57, no. 4, p. 045013, 2020.
- [22] S. Li, F. Bielsa, A. Kiss, and H. Fang, "Self-attraction mapping and an update on local gravitational acceleration measurement in BIPM kibble balance," *Metrologia*, vol. 54, no. 4, p. 445, 2017.
- [23] S. Li, W. Zhao, S. Huang, and X. Yu, "Design of the Tsinghua openhardware tabletop Kibble balance," in *2022 Conference on Precision Electromagnetic Measurements (CPEM 2022)*, 2022, pp. 1–2.
- [24] S. Li, Y. Ma, W. Zhao, S. Huang, and X. Yu, "Design of the Tsinghua tabletop Kibble balance," *IEEE Transactions on Instrumentation and Measurement*, 2023.
- [25] T. Niebauer, G. Sasagawa, J. E. Faller, R. Hilt, and F. Klotz, "A new generation of absolute gravimeters," *Metrologia*, vol. 32, no. 3, p. 159, 1995.
- [26] A. Peters, K. Y. Chung, and S. Chu, "Measurement of gravitational acceleration by dropping atoms," *Nature*, vol. 400, no. 6747, pp. 849–852, 1999.
- [27] M. Van Camp and P. Vauterin, "Tsoft: graphical and interactive software for the analysis of time series and earth tides," *Computers & Geosciences*, vol. 31, no. 5, pp. 631–640, 2005.
- [28] H.-G. Wenzel, "The nanogal software: Earth tide data processing package ETERNA 3.30," *Bull. Inf. Marées Terrestres*, vol. 124, pp. 9425–9439, 1996.
- [29] T. Hartmann and H.-G. Wenzel, "The HW95 tidal potential catalogue," *Geophysical research letters*, vol. 22, no. 24, pp. 3553–3556, 1995.
- [30] M. Stock, P. Conceição, H. Fang, F. Bielsa, A. Kiss, L. Nielsen, F. Beaudoux, P. Espel, M. Thomas, D. Ziane *et al.*, "Final report on the CCM key comparison of kilogram realizations CCM. M-K8. 2021," *Metrologia*, vol. 60, no. 1A, p. 07003, 2023.
- [31] O. Francis, "Performance assessment of the relative gravimeter Scintrex CG-6," *Journal of Geodesy*, vol. 95, no. 116, 2021.
- [32] Warburton, Richard J. *et al.*, "Initial Results with the New GWR iGrav™ Superconducting Gravity Meter," 2010.
- [33] U. S. Atmosphere, *US standard atmosphere*. National Oceanic and Atmospheric Administration, 1976.